




NOTES:

1. SOLID BASE ARRAYS ARE NOT SELF COOLING AND REQUIRE THE USE OF A THERMAL TRANSFER DECK (TTD) WITH A CHILLER OR A SIMILAR DEVICE. THEY MAY ALSO BE PLACED IN A COOLING BAY.

REVISION HISTORY			
REV	DESCRIPTION	DATE	APPROVED
1A	INITIAL RELEASE	7/29/2025	ENGINEERING



DRAWN SAS email engineering@analytical-sales.com	7/29/2025	ANALYTICAL SALES & SERVICES, INC © 2025		
		TITLE		
		24-18_lumidoxII_solid_base_rev1a_PUBLIC		
		SIZE		REV
		B		1A
		SCALE	1 : 1	SHEET 1 OF 1